

Certificate



Certificate no.

CU 72082555 01

License Holder:

XEI Scientific
1755 East Bayshore Road Suite 17
Redwood City CA 94063
USA

Manufacturing Plant:

XEI Scientific
1755 East Bayshore Road Suite 17
Redwood City CA 94063
USA

Test report no.: USA-JL 30372945 009

Client Reference: Ron Vane

Tested to: UL 61010-1:2004
CAN/CSA-C22.2 61010-1:2004

Certified Product: RF Plasma Cleaning System

License Fee - Units

Model Designation: Evactron C, 25, 40, 45

7

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 150W

Protection Class: I

7

Appendix: 1, 1-2

Licensed Test mark:

Signature

Date of Issue

(day/mo/yr)

07/11/2008



Martin Magla
Dipl.-Ing. M. Magla
QA Certification Officer