

Date : 10/06/2015

XEI Scientific
1755 East Bayshore Road Suite 17
Redwood City CA 94063
USA

Attn: Michael Cable

Re. : US US Certificate

Type of Equipment : RF Plasma Cleaning System
Model Designation : See Certificate
Certificate No. : US 72152386 0001
File No. : 30372945 022
Engineer/Contact : Calvin Lee
Standard(s) : UL 61010-1:2012

Dear Mr. Cable,


The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. US 72152386 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. M. Glagla
QA Certification Officer

Enclosure

TUV Rheinland
of North America, Inc.
North American Headquarters

12 Commerce Road
Newtown, CT 06470

Tel 203-426-0888
Fax 203-426-4009
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Certificate



Certificate no.

US 72152386 01

License Holder:

XEI Scientific
1755 East Bayshore Road Suite 17
Redwood City CA 94063
USA

Manufacturing Plant:

XEI Scientific
1755 East Bayshore Road Suite 17
Redwood City CA 94063
USA

Test report no.: USA-CAL 30372945 022

Client Reference: Michael Cable

Tested to: UL 61010-1:2012

Certified Product: RF Plasma Cleaning System

License Fee - Units

Model Designation:

Evactron C, Evactron 25, Evactron 40,
Evactron 45, Evactron 25Z, Evactron 40Z,
Evactron 45Z, Evactron EP, Evactron ES

7

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 150W

Protection Class: I

Special Remarks: Replaces Certificate CU72082555.

7

Appendix: 1, 1-2

Licensed Test mark:



Date of Issue

(day/mo/yr)

05/10/2015